

SmartCast®

ENSURE THE PATIENT IS CENTERED ON THE CHAIR (rev. 6/20)

If a patient sits too close to the sides of the chair, the patient and/or the SmartCast Foot Positioning System® can lean due to compression of the chair's cushioning. This can cause the leg/knee to rotate outward/externally, which is not desirable.

The patient's knee cap should always be pointing straight up. When the patient initially sits on the chair, take a moment to check their positioning.

