

SmartCast®

EXTEND THE FOOT 3" BEYOND THE END OF THE TABLE/CHAIR

Fully capturing the back of the patient's heel is a crucial part of the scanning process. The iPad/Structure Sensor can only scan what it can "see" with its camera and sensor.

If the foot is extended LESS than 3" beyond the end of the chair, it's impossible to fully capture the heel. A bit more than 3" is okay, although if the foot extends too far, the SmartCast Foot Positioning System® can become unstable and difficult to align.

