



SmartCast®

USE THE HEEL TAB FOR INITIAL POSITIONING ONLY (rev. 6/20)

The Heel Tab should touch the patient's heel during initial positioning of the patient's foot, but should be rotated away from the heel prior to scanning the foot.

If not moved, the Heel Tab can slightly deform the patient's heel while adding extra data to the scan, making processing more challenging for our technicians.

If a patient is scanned with the Heel Tab touching the foot, move the Heel Tab and rescan the patient.

