

SmartCast®

ENSURE THE PATIENT'S LEGS ARE ALWAYS ON TOP OF THE CHAIR (rev. 6/20)

The leg *not* being positioned/aligned with the SmartCast Foot Positioning System® should be ON the chair, straight ahead and fully extended (not bent at the knee).

Patients tend to let their legs dangle from the side of the chair, which negatively impacts alignment.

Correct:

